



PuraPly® Antimicrobial XT Wound Matrix

Description

PuraPly Antimicrobial XT Wound Matrix is a cross-linked extracellular matrix (ECM) scaffolding that consists of a collagen sheet coated with 0.1% polyhexamethylenebiguanide hydrochloride (PHMB) intended for the management of wounds. PuraPly Antimicrobial XT Wound Matrix is supplied dry in sheet form. The device is packaged in sterile, sealed single pouches.

Indications

PuraPly Antimicrobial XT Wound Matrix is indicated, for the management of wounds and as an effective barrier to resist microbial colonization within the device and reduce microbes penetrating through the device.

PuraPly Antimicrobial XT Wound Matrix is indicated for the management of:

- Partial and full-thickness wounds
- Pressure ulcers
- Venous ulcers
- Diabetic ulcers
- Chronic vascular ulcers
- Tunneled/undermined wounds
- Surgical wounds (donor sites/grafts, post-Mohs' surgery, post-laser surgery, podiatric, wound dehiscence)
- Trauma wounds (abrasions, lacerations, second-degree burns, skin tears)
- Draining wounds

This device is intended for one-time use. The device is intended for single patient use only. Do not reuse.

Contraindications

- PuraPly Antimicrobial XT Wound Matrix is derived from a porcine source and should not be used in patients with known sensitivity to porcine materials.
- PuraPly Antimicrobial XT Wound Matrix is not indicated for use in third degree burns.
- Do not use on individuals with a known sensitivity to polyhexamethylenebiguanide hydrochloride (PHMB)

Precautions

- **Do not resterilize.** Discard all open and unused portions of PuraPly Antimicrobial XT Wound Matrix.
- PuraPly Antimicrobial XT Wound Matrix is sterile if the package is dry, unopened and undamaged. Do not use if the package seal is broken.
- PuraPly Antimicrobial XT Wound Matrix must be used prior to the expiration date.
- Discard PuraPly Antimicrobial XT Wound Matrix if mishandling has caused possible damage or contamination.

- PuraPly Antimicrobial XT Wound Matrix should not be applied until excessive exudates, bleeding, acute infection and significant swelling are controlled.

Potential Complications

The following are possible with the use of PuraPly Antimicrobial XT Wound Matrix. If any of these conditions occur, PuraPly Antimicrobial XT Wound Matrix should be removed.

- Worsening Infection
- Chronic inflammation (Initial application of PuraPly Antimicrobial XT Wound Matrix may be associated with transient, mild, localized inflammation)
- Allergic reaction
- Excessive redness, pain, swelling or blistering

Instructions for Use

Note: These recommendations are designed to serve only as a general guideline. They are not intended to supersede institutional protocols or professional clinical judgment concerning patient care.

CAUTION: Federal (U.S.A) law restricts this device to sale by or on the order of a physician or properly licensed practitioner.

PuraPly Antimicrobial XT Wound Matrix can be applied from the onset and for the duration of the wound, with subsequent application at the discretion of the health care practitioner. PuraPly Antimicrobial XT

Wound Matrix will naturally be broken down and resorbed into the wound and is not intended to be removed.

Always handle PuraPly Antimicrobial XT Wound Matrix using aseptic technique.

1. Prepare wound area using standard methods to ensure wound is free of debris and necrotic tissue. If necessary, surgically debride the wound to ensure the wound edges contain viable tissue.
2. To apply, cut PuraPly Antimicrobial XT Wound Matrix to the outline of the wound area. If the wound is larger than a single PuraPly Antimicrobial XT Wound Matrix, then multiple PuraPly Antimicrobial XT Wound Matrix may be used. Overlap adjoining PuraPly Antimicrobial XT Wound Matrix to provide coverage of the entire wound. For ease of handling apply PuraPly Antimicrobial XT Wound Matrix by placing in a dry state over the wound and rehydrate the sheet using sterile saline or other isotonic solution.
3. Place the edge of PuraPly Antimicrobial XT Wound Matrix in contact with the intact tissue. Smooth PuraPly Antimicrobial XT Wound Matrix into place to ensure it is in contact with the underlying wound bed. Anchor PuraPly Antimicrobial XT Wound Matrix using preferred fixation method such as sutures or adhesive strips. The optimal fixation method is determined by wound location, size, depth and user preference.
4. Apply a non-adherent dressing followed by secondary dressing as appropriate for the type and stage of wound.

NOTE: If excess exudate collects under PuraPly Antimicrobial XT Wound Matrix, small openings can be cut in PuraPly Antimicrobial XT Wound Matrix to allow the exudate to drain.

IMPORTANT: After application, use an appropriate, non-adherent dressing to maintain a moist wound environment. The optimal secondary dressing is determined by wound location, size, depth and user preference. Change the secondary dressing as needed to maintain a moist, clean wound area. Take care to avoid removing PuraPly Antimicrobial XT Wound Matrix when the secondary dressing is changed. Frequency of secondary dressing change will be dependent upon volume of exudate produced and type of dressing used. PuraPly Antimicrobial XT Wound Matrix is not intended to be removed. PuraPly® Antimicrobial XT Wound Matrix will gradually resorb and may form into a caramel-colored gel. As healing occurs, do not forcibly remove PuraPly Antimicrobial XT Wound Matrix that adheres to the wound. The wound should be reevaluated on a weekly basis. Subsequent application of PuraPly Antimicrobial XT Wound Matrix is at the discretion of the health care practitioner.

Storage


Do not freeze or expose to excessive heat.


Sterilization

PuraPly Antimicrobial XT Wound Matrix has been sterilized with gamma irradiation.

Symbols Used on Labeling

 Sterilized using irradiation

 Use by Date

 Consult Instructions for use

 Batch Code

Manufactured and Distributed by:

 Organogenesis inc.
Empowering Healing

Organogenesis Inc.
Canton MA, 02021 USA
1-888-432-5232
www.organogenesis.com

Revision Date: August 2023
Pat 7,604,816, exp 10/08/2026